

shall be convenient to and have full access to the facility and, shall provide the inspector both visual and acoustic privacy.

(2) For a site with a single fuel facility licensed pursuant to part 70, the space provided shall be adequate to accommodate a full-time inspector, a part-time secretary and transient NRC personnel and will be generally commensurate with other office facilities at the site. A space of 250 square feet either within the site's office complex or in an office trailer or other on site space is suggested as a guide. For sites containing multiple fuel facilities, additional space may be requested to accommodate additional full-time inspector(s). The office space that is provided shall be subject to the approval of the Director, Office of Nuclear Material Safety and Safeguards or the appropriate NRC Regional Administrator. All furniture, supplies and communication equipment will be furnished by the Commission.

(3) The licensee shall afford any NRC resident inspector assigned to that site or other NRC inspectors identified by the Director, Office of Nuclear Material Safety and Safeguards, as likely to inspect the facility, immediate unfettered access, equivalent to access provided regular plant employees, following proper identification and compliance with applicable access control measures for security, radiological protection, and personal safety.

[21 FR 764, Feb. 3, 1956. Redesignated at 25 FR 1607, Feb. 25, 1960, and 25 FR 12730, Dec. 13, 1960, and amended at 32 FR 2563, Feb. 7, 1967; 44 FR 47919, Aug. 16, 1979; 52 FR 31612, Aug. 21, 1987; 54 FR 6877, Feb. 15, 1989; 55 FR 5979, Feb. 21, 1990]

§ 70.56 Tests.

Each licensee shall perform, or permit the Commission to perform, such tests as the Commission deems appropriate or necessary for the administration of the regulations in this part, including tests of (a) special nuclear material, (b) facilities wherein special nuclear material is utilized, produced or stored, (c) radiation detection and monitoring instruments, and (d) other equipment and devices used in connec-

tion with the production, utilization or storage of special nuclear material.

[21 FR 764, Feb. 3, 1956. Redesignated at 25 FR 1607, Feb. 25, 1960, and 25 FR 12730, Dec. 13, 1960]

§ 70.57 Measurement control program for special nuclear materials control and accounting.

(a) As used in this section:

(1) *Measurement* includes sampling and means the determination of mass, volume, quantity, composition or other property of a material where such determinations are used for special nuclear material control and accounting purposes.

(2) *Measurement system* means all of the apparatus, equipment, instruments and procedures used in performing a measurement.

(3) *Reference standard* means a material, device, or instrument whose assigned value is known relative to national standards or nationally accepted measurement systems.

(4) *Traceability* means the ability to relate individual measurement results to national standards or nationally accepted measurement systems through an unbroken chain of comparisons.

(5) *Random error* refers to the variation encountered in all measurement work, characterized by the random occurrence of both positive and negative deviations from a mean value.

(6) A *systematic error* is a constant unidirectional component of error that affects all members of a data set; its value can, in some instances, be estimated by the deviation of the mean of a measurement process from a reference value. A systematic error whose value has been determined in this manner is called a bias, whose effect can be corrected for.

(7) *Uncertainty* is the extent to which a measurement result is in doubt because of the effects of random error variances and the limits of systematic errors associated with a measurement process, after the measurements result has been corrected for bias.

(8) *Calibration* means the process of determining the numerical relationship between the observed output of a measurement system and the value, based upon reference standards, of the characteristics being measured.